

Boule Con-Diff

CONTROL

IVD



| Control | Lot | |
|-----------|----------|------------|
| Tri-Level | 22210-1K | 2023-03-02 |
| Low | 22210-11 | 2023-03-02 |
| Normal | 22210-12 | 2023-03-02 |
| High | 22210-13 | 2023-03-06 |

Open vial stability 14 days ¹

| SWELAB | CONTROL | | L |
|-----------------------|---------------------|-------------------|----------------------|
| | LOT | | 22210-11 |
| Swelab Alfa/Alfa Plus | | Mean ² | ± Range ³ |
| RBC | 10 ¹² /L | 2.25 | 0.12 |
| MCV | fL | 66.2 | 5.0 |
| RDW | % | 14.6 | 4.0 |
| Hct | % | 14.9 | 2.5 |
| Plt | 10 ⁹ /L | 75 | 15 |
| MPV | fL | 10.8 | 1.5 |
| WBC | 10 ⁹ /L | 3.5 | 0.3 |
| Hgb | g/dL | 5.4 | 0.3 |
| MCH | pg | 24.0 | 1.5 |
| MCHC | g/dL | 36.2 | 3.0 |
| Lymph# | 10 ⁹ /L | 1.6 | 0.4 |
| Mid# | 10 ⁹ /L | 0.1 | 0.3 |
| Gran# | 10 ⁹ /L | 1.8 | 0.8 |
| Lymph | % | 46.1 | 7.0 |
| Mid | % | 3.8 | 5.0 |
| Gran | % | 50.1 | 7.0 |

| CONTROL | | N |
|---------|-------------------|----------------------|
| LOT | | 22210-12 |
| | Mean ² | ± Range ³ |
| | 4.20 | 0.18 |
| | 86.6 | 5.0 |
| | 12.3 | 4.0 |
| | 36.4 | 3.0 |
| | 235 | 30 |
| | 10.9 | 1.5 |
| | 8.8 | 0.6 |
| | 12.6 | 0.4 |
| | 30.0 | 1.5 |
| | 34.6 | 3.0 |
| | 4.3 | 0.6 |
| | 0.4 | 0.5 |
| | 4.1 | 1.0 |
| | 49.0 | 5.0 |
| | 4.1 | 5.0 |
| | 46.9 | 5.0 |

| CONTROL | | H |
|---------|-------------------|----------------------|
| LOT | | 22210-13 |
| | Mean ² | ± Range ³ |
| | 5.07 | 0.22 |
| | 94.3 | 5.0 |
| | 12.3 | 4.0 |
| | 47.8 | 3.5 |
| | 544 | 60 |
| | 11.2 | 1.5 |
| | 21.1 | 1.8 |
| | 16.2 | 0.5 |
| | 32.0 | 1.5 |
| | 33.9 | 3.0 |
| | 13.2 | 1.5 |
| | 0.7 | 1.0 |
| | 7.2 | 2.5 |
| | 62.4 | 5.0 |
| | 3.3 | 5.0 |
| | 34.3 | 5.0 |

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfacute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Rasjon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22210-11
Prod.: Con-Diff Low
ID: 2221011+
Exp.: 2-Mar-2023

Instr.: Swelab Alfa
Lot: 22210-12
Prod.: Con-Diff Norm
ID: 2221012+
Exp.: 2-Mar-2023

Instr.: Swelab Alfa
Lot: 22210-13
Prod.: Con-Diff High
ID: 2221013+
Exp.: 6-Mar-2023

